H&A-107

IN THE CLAIMS

(Currently Amended) A therapeutic ultrasound system,
 comprising:

an ultrasonic transducer which irradiates a therapeutic ultrasound on a region to be treated coagulated by heat of said therapeutic ultrasound;

an input unit to which information of irradiation of said therapeutic ultrasound is inputted, said information including a continuous insonation time of said therapeutic ultrasound;

a control unit to which said information is inputted and which controls irradiation of said therapeutic ultrasound;

a sound detector which detects, during exposure of said therapeutic ultrasound, an audible sound generated in said region to be treated coagulated;

a waveform analyzing unit which obtains a crosscorrelation function between a waveform of the detected
audible sound and a typical waveform of an audible sound
previously obtained, during irradiation of said therapeutic
ultrasound, in a region to be treated coagulated; and

a unit which detects a point of time of detection of the audible sound using said cross-correlation function, and which sends a signal expressing detection of the audible sound to said control unit,

2

BEST AVAILABLE COPY

H&A-107

wherein said ultrasonic transducer is controlled by said control unit to irradiate irradiates said therapeutic ultrasound on said region to be treated coagulated, in said continuous insonation time from said point of time of detection of the audible sound by said sound detector until expiration of said continuous insonation time.

2-3. (Canceled).

4. (Currently Amended) A therapeutic ultrasound system, comprising:

an ultrasonic transducer which irradiates a therapeutic ultrasound on a region to be treated coagulated by heat of said therapeutic ultrasound;

an input unit to which [[an]] information of irradiation of said therapeutic ultrasound is inputted, said information including a continuous insonation time of said therapeutic ultrasound;

a control unit to which said information is inputted and which controls irradiation of said therapeutic ultrasound;

a sound detector which detects, during exposure of said therapeutic ultrasound, an audible sound generated in said region to be treated coagulated;

H&A-107

a signal processing unit which obtains a FFT spectrum of said audible sound before a start of irradiation of said therapeutic ultrasound and a FFT spectrum of said audible sound after the start of irradiation of said therapeutic ultrasound; and

a unit which detects a point of time of detection of the audible sound by comparing the FFT spectrum before the start of irradiation of said therapeutic ultrasound with the FFT spectrum after the start of irradiation of said therapeutic ultrasound.

wherein said ultrasonic transducer is controlled by said control unit to irradiate irradiates said therapeutic ultrasound on said region to be treated coagulated, in said continuous insonation time from said point of time of detection of the audible sound by said sound detector until expiration of said continuous insonation time.

5. (Canceled)

6. (Previously presented) A therapeutic ultrasound system according to Claim 4, wherein said unit calculates a first integrated value of the FFT spectrum before the start of irradiation of said therapeutic ultrasound in a preset

H&A-107

frequency interval and a second integrated value of the FFT spectrum after the start of irradiation of said therapeutic ultrasound in said preset frequency interval, and detects said point of time of detection of the audible sound by comparing said first integrated value with said second integrated value.

- 7. (Previously presented) A therapeutic ultrasound system according to Claim 6, wherein said preset frequency interval is 250 Hz to 550 Hz.
- 8. (Previously presented) A therapeutic ultrasound system according to Claim 4, wherein said unit detects said point of time of detection of the audible sound by comparing a signal level at a particular frequency of the FFT spectrum before the start of irradiation of said therapeutic ultrasound with a signal level at said particular frequency of the FFT spectrum after the start of irradiation of said therapeutic ultrasound.

H&A-107

9. (Currently amended) A therapeutic ultrasound system, comprising:

an ultrasonic transducer which irradiates a therapeutic ultrasound on a region to be treated coagulated by heat of said therapeutic ultrasound;

an input unit to which [[an]] information of irradiation of said therapeutic ultrasound is inputted, said information including a continuous insonation time of said therapeutic ultrasound;

a control unit to which said information s inputted and which controls irradiation of said therapeutic ultrasound;

a sound detector which detects, during exposure of said therapeutic ultrasound, an audible sound generated in said region to be treated coagulated;

a signal processing unit which obtains a FFT spectrum of the detected audible sound; and

a waveform analyzing unit which obtains a crosscorrelation function between the FFT spectrum of the detected
audible sound and a typical FFT spectrum of a typical waveform
of an audible sound previously obtained, during irradiation of
said therapeutic ultrasound, in a region to be treated
coagulated; and

H&A-107

a unit which detects a point of time of detection of the audible sound using said cross-correlation function and sends a signal expressing detection of the audible sound to said control unit,

wherein said ultrasonic transducer is controlled by said control unit to irradiate irradiates said therapeutic ultrasound on said region to be treated coagulated, in said continuous insonation time from said point of time of detection of the audible sound by said sound detector until expiration of said continuous insonation time.

10. (New) The therapeutic ultrasound system according to Claim 1, wherein said sound detector detects audible sound in the range of 200 to 900 Hz.

This Page is Inserted by IFW Indexing and Scanning Operations and is not part of the Official Record

BEST AVAILABLE IMAGES

Defective images within this document are accurate representations of the original documents submitted by the applicant.

Defects in the images include but are not limited to the items checked:

BLACK BORDERS

IMAGE CUT OFF AT TOP, BOTTOM OR SIDES

FADED TEXT OR DRAWING

BLURRED OR ILLEGIBLE TEXT OR DRAWING

SKEWED/SLANTED IMAGES

COLOR OR BLACK AND WHITE PHOTOGRAPHS

GRAY SCALE DOCUMENTS

LINES OR MARKS ON ORIGINAL DOCUMENT

REFERENCE(S) OR EXHIBIT(S) SUBMITTED ARE POOR QUALITY

IMAGES ARE BEST AVAILABLE COPY.

☐ OTHER: _

As rescanning these documents will not correct the image problems checked, please do not report these problems to the IFW Image Problem Mailbox.